



ADMHE

ENGLISH TEXT

INQUIRY No : 407301

SUPPLEMENT No 1

DATE : 08.02.2013

OBJECT : PORTABLE DEVICE FOR MEASURING POWER TRANSFORMERS WINDING RESISTANCE

1. This Supplement No 1, of the above mentioned Inquiry, is issued to replace the paragraphs 10, 11, 12, 16 and 18 of the Technical Specification as follows :
 - Paragraph 10 will be replaced as follows :
10. Measuring ranges in steps with nominal values from 1mOhm to 100Ohms (automatically) with automatic measurement compensation in 20°C.
 - Paragraph 11 will be replaced as follows :
11. Minimum accuracy : ± 0.25 %.
 - Paragraph 12 will be replaced as follows :
12. Test current at least 50 A DC.
 - The subparagraph γ , is removed, given that the instrument will have the possibility of specimen full discharge when test is completed.
 - Paragraph 18 will be replaced as follows :
18. The supplier has to provide together with the instrument a complete operation instruction book in Greek of English.
2. All Terms and Conditions of the Initial Inquiry, except as expressly modified by this Supplement No 1, will remain in full force and effect.